

June 30, 2021

Dräger Medical Verification of Aruba AP335 and 70xx Controller for OneNet® (shared Dräger Infinity network)

Dear Madam/Sir,

Dräger Medical Systems is pleased to announce the verification of the Aruba Access Point AP335 with 70xx Controllers running 8.6.0.9 software. These components have been extended to the Infinity OneNet approved network equipment checklist, OneNet is the award-winning solution introduced by Dräger Medical that enables the integration of Infinity patient monitors as part of a single hospital wide network.

The following network components are verified for use with the Dräger Medical Infinity Network:

WLAN Controller	Access Points	Tested Software Version	Limitations
Aruba 7008 Controller	Aruba Wireless AP-335	8.6.0.9	M300 HW versions 0.4 or later supported

The goal of this verification initiative is to investigate situations where customers want to operate Dräger and non-Dräger clients on shared IEEE 802.11 a/b/g/n/ac/ax access points. The above referenced access point has passed the simulated tests where an access points were loaded with eight wireless Infinity Delta patient monitors, 16 Infinity M300 monitors, eight M540 standalone monitors, as well as competing traffic.

Please be advised it remains the customer's responsibility to actively monitor potential sources of wireless interference as well as changes in the wireless traffic load.

Yours Sincerely,



Tom Kinahan
Sr. Product Manager
Monitoring, Systems & IT
Tel: +1 978 379-6565
tom.kinahan@draeger.com

Draeger Medical Systems, Inc.

6 Tech Drive
Andover, MA 01810 USA
Phone: (978) 379-8000
www.draeger.com