

December 16, 2016

Dräger Medical Verification of Aruba 72xx Mobility Controller, 70xx Cloud Services Controller and Aruba AP-204, AP-205, AP-205-H, AP-214, AP-215, AP-224, AP-225, AP-228, AP-274, AP-275, AP-277 Access Points with Wireless Network for OneNet® (shared Dräger Infinity network) Installations

Dear Madam/Sir,

Dräger Medical is pleased to announce the verification of the Aruba 72xx Mobility Controller, 70xx Cloud Services Controller and Aruba AP-204, AP-205, AP-205-H, AP-214, AP-215, AP-224, AP-225, AP-228, AP-274, AP-275, AP-277 Access Points with 6.5.0.0 firmware. These components have been extended to the Infinity OneNet approved network equipment checklist. Infinity OneNet is the award-winning solution introduced by Dräger Medical that enables the integration of Infinity patient monitors as part of a single hospital wide network.

The following network components are verified for use with the Dräger Medical OneNet network:

WLAN Controller	Access Points	Tested Software Version
Aruba 72xx Mobility Controller 70xx Cloud Services Controller	AP-224 AP-204 AP-225 AP-205 AP-228 AP-205-H AP-274 AP-275 AP-214 AP-277 AP-215	6.5.0.0

The goal of this verification initiative is to investigate situations where customers want to operate Dräger and non-Dräger clients on shared IEEE 802.11 a/b/g/n access points. The above referenced Aruba wireless controller and access points passed the simulated tests where an access points was loaded with 6 wireless Infinity Delta patient monitors, 12 Infinity M300 monitors, 2 IACS patient monitors, and 12 simulated wired patient monitors as well as competing traffic.

Please be advised it remains the customer's responsibility to actively monitor potential sources of wireless interference as well as changes in the wireless traffic load.

Yours Sincerely,



Angela Noble
Product Manager, Infinity OneNet
Monitoring, Systems & IT
Tel: +1 978 379-8514
angela.noble@draeger.com

Draeger Medical Systems, Inc.

6 Tech Drive
Andover, MA 01810 USA
Phone: (978) 379-8000
www.draeger.com